



**DEPARTMENT OF CHEMICAL ENGINEERING**  
**UNIVERSITY OF CALCUTTA**

**NOTICE**

Ref: no. CHE-ENR/WB DST-FIST/HOD/02/2018-19 date 7.06.18

Quotations are invited within fifteen (15) days from the date of notice, in duplicate, from vendors for purchase of Stationery/ Electrical Goods/ Chemicals / Glass Goods / **Lab. Instrument & Equipment**/ Wooden & Steel Furniture/ Computer & Accessories/ A.C Machine, Refrigerator & Similar other Machines. (tick the applicable one) The quotations should accompany copies of valid trade license, GST certificates as applicable, PAN and other relevant documents as per University Purchase Rules

Faculty member

Head of the Department

*Prof. & Head*  
Dept. of Chemical Engineering  
University of Calcutta  
A. P. C. Road,  
Kolkata - 700 009

**List and specification of the items required**

**Analytical Balance:**

Analytical Balance, 120 g x 0.1 mg. Three- door glass draft shield. Weighing modes include parts counting percent weighing and check weighing. Sealed keypad, splash proof design and removal stainless steel weighing pan. Pan size: 90 mm. Stabilization time: Fast stabilization and response time. Power: A/C Adaptor 220 V. Calibration: Internal or External with calibration mass. Display: backlit LCD should provide clear visibility. USB interface for connection to computer